



財團法人全國認證基金會
Taiwan Accreditation Foundation

認 證 證 書

(證書編號：L1451-240402)

茲證明

財團法人國家實驗研究院
國家地震工程研究中心地震模擬實驗室

台北市辛亥路三段 200 號

為本會認證之實驗室

認 證 依 據：ISO/IEC 17025：2017；CNS 17025：2018

認 證 編 號：1451

初 次 認 證 日 期：九十四年五月十七日

認 證 有 效 期 間：一百一十三年四月二日至一百一十六年四月一日止

認 證 範 圍：測試領域，如續頁

董事長

陳怡鈴



掃描確認真偽

中華民國一一三年四月二日

認證編號：1451

實驗室主管：曾建創

■ 21.99 建築材料

機械元件

A002 振動測試

(1) AC 156 (2010, 2020 年 12 月修訂) Sec. 6.4.1, 6.4.5, 6.5.1, 6.5.3

(2) IEEE Std 693-2018 Sec. A.1.2.5 排除 a) 與 b), A.1.2.6, 圖 A.2

(3) GR 63 CORE:2017 Sec. 4.4.1.2/R4-66、R4-67、5.4.1、5.4.1.2、5.4.1.3/Frame- Level、
Frame Level Instrumentation Configuration/1, 3

振動台頻率範圍: (1 to 50) Hz

全載最大荷重: 392.266 kN, 40000 kgf

報告簽署人: 吳紀弘; 鄭橙標

(以下空白)





財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L1451-240402)

This is to certify that

**National Applied Research Laboratories
Seismic Simulation Laboratory, National Center for Research on
Earthquake Engineering**

No.200, Sec.3, Hsinhai Rd., Taipei City 106, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 1451

Originally Accredited : May 17, 2005

Effective Period : April 02, 2024 to April 01, 2027

Accredited Scope : Testing Field, see described in the Appendix



Scan to verify

Yi-Ling Chen

Yi-Ling Chen
President, Taiwan Accreditation Foundation
April 02, 2024

Accreditation Number : 1451

Laboratory Head : TSENG, Chien-Chuang

■ 21. 99 Building Materials

Machine Devices

A002 Vibration Testing

(1) AC 156 (2010, Editorially revised December 2020) Sec. 6.4.1, 6.4.5, 6.5.1, 6.5.3

(2) IEEE Std 693-2018 Sec. A.1.2.5 exclude a) and b), A.1.2.6, Fig A.2

(3) GR 63 CORE: 2017 Sec. 4.4.1.2/R4-66, R4-67, 5.4.1, 5.4.1.2, 5.4.1.3/Frame-Level 、
Frame Level Instrumentation Configuration/1, 3

The Shaking Table Frequency Range: (1 to 50) Hz

Maximum Payload: 392.266 kN, 40000 kgf

Approval Signatory: WU, Chi-Hung; CHENG, Cheng-Piao

(Null below)

