



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L1451-250328)

This is to certify that

National Institutes of Applied Research
Seismic Simulation Laboratory, National Center for Research on
Earthquake Engineering
No.200, Sec.3, Hsinhai Rd., Taipei City 106, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 1451

Originally Accredited : May 17, 2005

Effective Period : April 02, 2024 to April 01, 2027

Accredited Scope : Testing Field, see described in the Appendix

Yi-Ling Chen



Scan to verify

Yi-Ling Chen
President, Taiwan Accreditation Foundation
March 28, 2025

Accreditation Number : 1451

Laboratory Head : TSENG, Chien-Chuang

▼ 21. 99 Building Materials

Machine Devices

A002 Vibration Testing

(1) AC 156 (2010, Editorially revised December 2020) Sec. 6.4.1, 6.4.5, 6.5.1, 6.5.3

(2) IEEE Std 693-2018 Sec. A.1.2.5 exclude a) and b) , A.1.2.6, Fig A.2

(3) GR 63 CORE: 2017 Sec. 4.4.1.2/R4-66, R4-67, 5.4.1, 5.4.1.2, 5.4.1.3/Frame- Level,

Frame Level Instrumentation Configuration /1, 3

The Shaking Table Frequency Range: (1 to 50) Hz

Maximum Payload: 392.266 kN, 40000 kgf

Approval Signatory: WU, Chi-Hung; CHENG, Cheng-Piao

(Null below)

